



CUPID
LIMITED

CupiKIT

Malaria PF- PAN Antigen Test Kit

For In Vitro Diagnostic Use Only

PRODUCT SPECIFICATION

- Chromatographic immunoassay
- Detection of antigen to Malaria PF-PAN
- Result interpretation time: 20 minutes
- Specimen: Whole Blood
- Sensitivity: 100 %
- Specificity: 100%
- Storage Temperature : 2°C -40°C

MATERIALS PROVIDED

PROVIDED

- Test Device
- Dropper
- Buffer Bottel
- Desiccant
- Instructions for Use

PACK SIZE: 50 Test Kit



**Manufactured by: Cupid Limited, A-68, M.I.D.C. (Malegaon),
Sinnar, Nashik - 422 113, Maharashtra, India,
Mfg. Lic. No.: MFG/IVD/2022/000065**